

METHODS & DESIGNS

A simple technique for lateralizing visual input that allows prolonged viewing

JAY J. MYERS and R. W. SPERRY

Division of Biology, California Institute of Technology, Pasadena, California 91125

A simplified technique is described for obtaining lateralization of visual input with prolonged viewing. This technique is based on the presence of constant normal lateral limits for horizontal rotation of the eyes with reference to the head. With head movement prevented by use of a standard bite bar, and the eyes rotated to the left and held at their lateral limit, the temporal half of the visual field of the left eye may be used for lateralized input to the right hemisphere or vice versa for input to the left hemisphere. Any form of visual stimuli or visually monitored task can be used if confined within one of the extreme temporal hemifields. In comparison to previous methods, this technique is technically simple, inexpensive, without significant risk or discomfort to the subject, readily applicable to normal and various brain-lesioned subjects, and permits prolonged in-depth viewing. An alternative version of this technique uses a stabilized spectacle frame fitted with adjustable central occluders set to allow vision through only one or both of the extreme temporal hemifields.

Techniques to selectively lateralize visual input to the separate cerebral hemispheres in human subjects must circumvent the natural tendency of eye movements to transfer the intended half-field stimuli across the vertical meridian into the unintended half-field and hemisphere. This has been achieved most commonly in the past by tachistoscopic methods that limit exposure of visual stimuli in the left or right hemifield to 150 msec or less. This restricts tachistoscopic testing to the use of relatively simple visual stimuli and thus excludes the use of many forms of tests for intelligence, memory, perception, emotion, and other cognitive functions that require more prolonged examination. The need for a better technique that allows prolonged viewing of more complex visual displays (e.g., sentences instead of single words, complex scenes and objects instead of simple line drawings, etc.) and lateralized viewing of manual performance has long been recognized, particularly for studies with commissurotomy subjects.

Early studies obtained prolonged lateralized exposure of visual stimuli by monitoring eye position with electro-oculograph (EOG) recording while stimuli were presented in the left or right hemifield (Butler & Norrsell, 1968; Trevarthen & Sperry, 1973). The stimulus could be removed or trial excluded whenever adverse eye movements were detected. Inherent inaccuracy due to artifact

and drift in the EOG potential has restricted this technique primarily to studies involving peripheral vision.

Special contact lenses which limit vision to a single hemifield or a portion thereof have also been employed to prolong visual lateralization (Dimond, Bureš, Farrington, & Brouwers, 1975; Zaidel, 1975). The sophisticated lens system developed by Zaidel is a variation of the stabilized image technique in which a half-field occluder, mounted on a collimator on a specially constructed scleral contact lens, is set in the focal plane of a small viewing lens and moves in unity with the eye. This has proved to be an important advance over previous techniques for many kinds of tests but is subject to a number of constraints that severely limit its application.

Another approach to prolonging lateralized visual input involves the use of a double-Purkinje image eye tracker in conjunction with a mechanical occluder or split-screen video display (Zaidel & Frazer, 1977). Presently in the process of development, this technique, if successful, should relieve some of the limitations of the contact lens approach in that there are no attachments to the eye and only a bite bar is fitted to each subject. However, the eye tracker appliance also is expensive and requires exacting technical adjustments that will presumably limit its use to relatively few laboratories.

In this paper we describe a comparatively simple, new technique for obtaining lateralized visual input with prolonged exposure of visual material. Successful in pilot tests and recent studies with commissurotomy patients, the new method offers a number of significant advantages over prior methods and opens new testing

This work was supported by U.S. Public Health Service Grant MH 3372 from the National Institute of Mental Health, National Research Service Award 5 T32 GM07737 from the National Institute of General Medical Sciences, the PEW Memorial Trust, and the F. P. Hixon Fund.

possibilities not available with previous techniques. In contrast to previous techniques, the new procedure is inexpensive, involves no significant risk or discomfort for the patient, imposes no stringent limits on the duration or frequency of testing, is widely applicable to normal or brain-damaged subjects, requires no advanced engineering or technological expertise either for its construction or its operation, and permits prolonged in-depth viewing under relatively natural conditions.

METHOD

The present technique is based upon the normal presence of constant lateral limits for horizontal rotation of the eyes with reference to the head. If head movement is prevented, and the eyes are rotated to the left and held at their lateral limit, the temporal half of the visual field of the left eye can be used for lateralizing input to the right hemisphere or vice versa for the left hemisphere. At these lateral limits of rotation when the head is stabilized, there is no means by which further eye movement can transfer stimuli in the extreme temporal testing fields across the midline into the unintended half-field. The concept was presented earlier in a brief abstract report (Sperry & Myers, 1981).

The technique can be readily adapted to normal and brain-damaged populations and can be coupled with reaction time responses to obtain laterality measures in the presence of the commissures. Any form of visual stimulus or test may be used, provided it is confined to one of the extreme temporal testing fields. Input lateralized to the separate hemispheres in this manner utilizes directly comparable pathways involving the nasal hemiretina and crossed optic pathway to each hemisphere. Since different eyes are used to lateralize input to the different hemispheres, differences between the eyes, such as differences in acuity, may need to be taken into account in making left-right comparisons. For subjects requiring visual correction, a set of corrective lenses may be laterally positioned to cover the testing fields.

No limit is imposed on the duration of lateralized viewing other than the natural development of ocular fatigue from holding the eyes in the extreme sideward position. If fatigue occurs, it is readily relieved by simply relaxing the eyes and directing the gaze forward or to the opposite side. In most cases the eyes need not be held at the lateral limits for an extended period. A series of short views may be allowed and their initiation left to the discretion of the subject. In our experience to date, a single viewing duration of up to approximately 10-15 sec appears to be well tolerated.

Head Fixation

A comfortable means for firmly holding the head of the subject in a fixed position is required during testing. A bite bar is quite adequate for this purpose. If interference with the clarity of oral response becomes a problem, the subject may be allowed to release the bar for response after removal of the stimuli. Other means of head fixation may be preferable in studies emphasizing oral performance concurrent with lateralized viewing. When freedom of head movement is important, it may be possible to substitute occluders affixed in spectacle fashion to a stabilized head frame and set to block vision in all but the extreme temporal testing fields.

The bite bar is made from 1/16-in. stainless steel cut into a 1/2-in.-wide U-shaped plate to fit the pattern of the bite. Perforations in the steel plate help to hold in place dental impression compound, which is heated and applied around both arms of the bar and then quickly resoftened in hot water for making the dental impression. Removable bridgework should be taken

out prior to fitting and before subsequent testing sessions. Numerous bars can be prepared in advance, ready to be reheated and fitted as needed. The bite bar plate is rigidly bolted to a solid supporting upright bar and firmly fixed to a table at a height comfortable for the subject when seated in testing position. The bite bar is mounted with a slight forward tilt of 4-5 deg to match the natural tilt of the jaw when the head is held level.

Determination of Testing Fields

The inner boundaries (i.e., the vertical midlines of the visual field at the limits of rotation) of the left and right testing hemifields (see Figure 1) are determined in advance for each subject for use in all subsequent testing. The lateral limits of fixation for most subjects are reported to be in the range of 45-50 deg from the forward midline (Duke-Elder & Wybar, 1973). Because of individual variation, we determine these limits for each subject by reference to the blind spot of each eye as described below. In the event of a superimposed visual field scotoma, as occurred with one of our subjects, these same measurements can be taken using the abnormal, rather than the normal, blind spot as the reference. In tests in which acuity is not critical and peripheral viewing is acceptable, the two testing fields can be safely demarcated with their inner boundaries at least 50 deg lateral from the forward midline, eliminating the need for more precise measurements. For expedience in making measurements and locating the testing field boundaries, a large perimetric semicircle is marked on the table top, centered below the middle of the bite bar, with a radius corresponding to the desired viewing distance and with angular displacements from the forward midline marked in the region of the testing fields.

The positions of the vertical midlines of the visual fields at the limits of lateral rotation are determined with the subject seated comfortably at the table and the bite bar in place. One eye is occluded while the subject fixates a small hairline cross presented at eye level, centered directly ahead on a white tangent screen at the desired viewing distance. The angular displacement from the vertical midline to the nasal edge of the blind spot, usually about 15 deg, is determined by moving a small black target of 1-deg diameter slowly back and forth across the screen along the horizontal meridian, noting on the screen where the subject either verbally or manually indicates its disappearance from view.

The subject is then directed to rotate the exposed eye laterally to its extreme limit while keeping the gaze level, and the screen is relocated to the approximate center of gaze. The location of the nasal edge of the blind spot with the eyes laterally rotated to their extreme limit is determined as before, by having the subject indicate the disappearance and reappearance of the target as it moves across the screen along the horizontal meridian. This is repeated until the subject is unable to cause the target




Figure 1. Schematic drawing of testing fields for lateralized visual input as set by lateral limits of eye rotation.

different techniques for lateralizing visual input. Results with the present technique replicate the early data obtained tachistoscopically (Gazzaniga & Sperry, 1967; Sperry & Gazzaniga, 1967; Sperry, Gazzaniga, & Bogen, 1969) in showing, for example, that the right hemisphere understands, at a moderately high level, words spoken aloud by the examiner and comprehends the meaning of printed object names exposed in the left visual field, as demonstrated by selective manual retrieval or by selection of the corresponding object or picture in a multiple-choice array. Comprehension of object names presented in the left visual hemifield is also apparent in the ability of these subjects to answer yes/no questions regarding characteristics of the object, such as its use, appearance, composition, likely context, and so forth, or to select the correct answer when alternatives are listed orally.

The present technique also replicates results obtained with the contact lens method (Zaidel, 1975, 1976, 1978) to the extent that these results have been tested. For example, the subjects can often correctly select from a multiple-choice array of pictures presented in the left testing field the item that matches a printed or spoken name or that fits an oral description, vocal cues, and so forth. The right hemisphere fails, however, to derive the sound of a word or picture name presented in the left testing field, nor can it analyze the word or name into component letters, as shown by the inability to make matches based on the sounds of words (rhyming, homonyms) or to select for component letters (e.g., which picture name starts with the letter "b"?). The right hemisphere is also seen with the present technique to be unable to decipher long, nonredundant sequences of spoken words, as when items of the Token test are performed in the left testing field.

Although tests with the scleral contact lens have brought results that in general confirm or extend, rather than revise, the earlier findings obtained by tachistoscopic methods, the two methods have yielded contradictory results in regard to the right hemisphere capacity for processing verbs as opposed to nouns. Zaidel (1976) reported equivalent oral comprehension by the right hemisphere for nouns vs. verbs and action names matched for frequency, whereas the early tachistoscopic findings indicated a selective deficiency in verb processing

by the right hemisphere. When we readministered the same set of verbs used in the test by Zaidel to L. B. and N. G. with the "lateral limits" technique, but used the infinitive form of the verbs rather than participles, the results showed a marked decline in the performance of the right hemisphere. These preliminary results thus favor the conclusion that the right hemisphere is relatively deficient in the processing of verbs as opposed to nouns.

REFERENCES

- BUTLER, S. R., & NORRSELL, U. Vocalization possibly initiated by the minor hemisphere. *Nature*, 1968, **220**, 793-794.
- DIMOND, S. J., BUREŠ, J., FARRINGTON, L. J., & BROUWERS, E. Y. M. The use of contact lenses for the lateralization of visual input in man. *Acta Psychologica*, 1975, **39**, 341-349.
- DUKE-ELDER, S., & WYBAR, K. Ocular motility and strabismus. In S. Duke-Elder (Ed.), *System of ophthalmology* (Vol. 6). St. Louis: C. V. Mosby, 1973.
- GAZZANIGA, M. S., & SPERRY, R. W. Language after section of the cerebral commissures. *Brain*, 1967, **90**, 131-148.
- SPERRY, R. W., & GAZZANIGA, M. S. Language following surgical disconnection of the hemispheres. In F. L. Darley (Ed.), *Brain mechanisms underlying speech and language*. New York: Grune & Stratton, 1967.
- SPERRY, R. W., GAZZANIGA, M. S., & BOGEN, J. E. Inter-hemispheric relationships: The neocortical commissures; syndromes of hemisphere disconnection. In P. J. Vinken & G. W. Bruyn (Eds.), *Handbook of clinical neurology*. Amsterdam: North-Holland, 1969.
- SPERRY, R. W., & MYERS, J. J. A simplified technique for lateralizing visual input. *Caltech Biology Annual Report*, 1981, 151-152. (Abstract)
- TREVARTHEN, C., & SPERRY, R. W. Perceptual unity of the ambient visual field in human commissurotomy patients. *Brain*, 1973, **96**, 547-570.
- ZAIDEL, E. A technique for presenting lateralized visual input with prolonged exposure. *Vision Research*, 1975, **15**, 283-289.
- ZAIDEL, E. Auditory vocabulary of the right hemisphere following brain bisection or hemidecortication. *Cortex*, 1976, **12**, 191-211.
- ZAIDEL, E. Lexical organization in the right hemisphere. In P. Buser & A. Rougeul-Buser (Eds.), *Cerebral correlates of conscious experience*. Amsterdam: Elsevier, 1978.
- ZAIDEL, E., & FRAZER, R. E. A universal half-field occluder for laterality research. *Caltech Biology Annual Report*, 1977, 137-138. (Abstract)

(Accepted for publication February 10, 1982.)